

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 10/11/2023

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:52
INTERNAL STANDARD	VERIFIED	03:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:52
BLANK TEST	0.00	03:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:53
BLANK TEST	0.00	03:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:55
BLANK TEST	0.00	03:56

