

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 04/06/2023

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:23
INTERNAL STANDARD	VERIFIED	05:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:23
BLANK TEST	0.00	05:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:24
BLANK TEST	0.00	05:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:25
BLANK TEST	0.00	05:26

