

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

*LEB*

Date: 12/13/2023

OPERATOR NAME:  
JOHN P CUSACK  
Permit #: 35563  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:18
INTERNAL STANDARD	VERIFIED	23:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:18
BLANK TEST	0.00	23:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:19
BLANK TEST	0.00	23:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:21
BLANK TEST	0.00	23:22

