

SOLUTION CHANGE

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 03/01/2024

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:03
INTERNAL STANDARD	VERIFIED	20:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:04
BLANK TEST	0.00	20:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:05
BLANK TEST	0.00	20:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:06
BLANK TEST	0.00	20:07

