

SOLUTION CHANGE *LE3*

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 01/01/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:47
INTERNAL STANDARD	VERIFIED	16:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	16:48
BLANK TEST	0.00	16:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	16:49
BLANK TEST	0.00	16:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	16:50
BLANK TEST	0.00	16:51

