

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 03/07/2024

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:28
INTERNAL STANDARD	VERIFIED	04:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:28
BLANK TEST	0.00	04:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:29
BLANK TEST	0.00	04:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:31
BLANK TEST	0.00	04:31

