

SUPERVISOR TEST

BFO

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 05/08/2024

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:44
INTERNAL STANDARD	VERIFIED	00:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:45
BLANK TEST	0.00	00:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:46
BLANK TEST	0.00	00:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:47
BLANK TEST	0.00	00:48

