

SUPERVISOR TEST

BFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 05/16/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:57
INTERNAL STANDARD	VERIFIED	22:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:57
BLANK TEST	0.00	22:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:59
BLANK TEST	0.00	22:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:00
BLANK TEST	0.00	23:01

