

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

BFO

Date: 10/09/2024

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:28
INTERNAL STANDARD	VERIFIED	22:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:28
BLANK TEST	0.00	22:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:29
BLANK TEST	0.00	22:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:31
BLANK TEST	0.00	22:32

