

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 11/11/2021

OPERATOR NAME:
JAMES P QUINN
Permit #: 49340
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:03
INTERNAL STANDARD	VERIFIED	04:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:04
BLANK TEST	0.00	04:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:05
BLANK TEST	0.00	04:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:06
BLANK TEST	0.00	04:07

CFO Datamaster
Weekly Test
11/11/21

