

SUPERVISOR TEST

CFO

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 01/27/2022

OPERATOR NAME:  
CHARLES C HAYNES  
Permit #: 32834  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:24
INTERNAL STANDARD	VERIFIED	20:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:24
BLANK TEST	0.00	20:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:26
BLANK TEST	0.00	20:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:27
BLANK TEST	0.00	20:28

