

**Infrared Instrument
Weekly Calibration Verification Test Record**

Department:

CFO

Instrument Serial #:

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
9-21-22	22190	34.01	34.01	34.01	MP4189	.10	.10	.10	Eilers	NO
9-28-22	22190	34.00	34.01	34.01	MP4189	.10	.10	.10	DANQUER	NO
10/01/22	22260	34.05	34.05	34.00	MP4189	.10	.10	.10	DANQUER	YES
10/5/22	22260	34.01	34.01	34.01	MP4189	.10	.10	.10	DANQUER	NO
10/12/22	22260	34.01	34.01	34.01	MP4189	.10	.10	.10	DIXON	NO
10/18/22	22260	34.01	34.01	34.00	MP4189	.10	.10	.10	DIXON	NO
10/24/22	22260	34.00	34.01	34.01	MP4189	.10	.10	.10	DIXON	NO
11-1-22	22260	34.01	34.01	34.01	MP4189	.10	.10	.10	Eilers	NO
11/2/22	22260	34.01	34.01	34.01	MP4189	.10	.10	.10	DANQUER	NO
11/9/22	22260	34.01	34.01	34.01	MP4189	.10	.10	.10	OSIKA	NO
11/16/22	22260	34.01	34.01	34.01	MP4189	.10	.10	.10	Eilers	NO
11/23/22	22260	34.01	34.01	34.00	MP4189	.10	.09	.10	DANQUER	NO

* - Please enter a ✓ to denote simulator solution change taking place