

SOLUTION CHANGE

CFO

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 03/01/2024

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:42
INTERNAL STANDARD	VERIFIED	21:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:42
BLANK TEST	0.00	21:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:44
BLANK TEST	0.00	21:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:45
BLANK TEST	0.00	21:46

