

SOLUTION CHANGE

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 03/21/2024

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:38
INTERNAL STANDARD	VERIFIED	22:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:38
BLANK TEST	0.00	22:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:39
BLANK TEST	0.00	22:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:41
BLANK TEST	0.00	22:42

