

SOLUTION CHANGE

CFO

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 07/01/2024

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:55
INTERNAL STANDARD	VERIFIED	19:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:55
BLANK TEST	0.00	19:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:57
BLANK TEST	0.00	19:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:58
BLANK TEST	0.00	19:59

