

SOLUTION CHANGE

CFD

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 01/01/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:35
INTERNAL STANDARD	VERIFIED	20:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:35
BLANK TEST	0.00	20:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:36
BLANK TEST	0.00	20:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:37
BLANK TEST	0.00	20:38

