

SUPERVISOR TEST

CFO

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 02/14/2024

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:25
INTERNAL STANDARD	VERIFIED	03:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:26
BLANK TEST	0.00	03:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:27
BLANK TEST	0.00	03:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:28
BLANK TEST	0.00	03:29

