

SUPERVISOR TEST

CFO

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 05/15/2024

OPERATOR NAME:  
BAILEY J DIXON  
Permit #: 50814  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:38
INTERNAL STANDARD	VERIFIED	17:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:39
BLANK TEST	0.00	17:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:40
BLANK TEST	0.00	17:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:41
BLANK TEST	0.00	17:42

