

SUPERVISOR TEST

CFD

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 05/30/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:36
INTERNAL STANDARD	VERIFIED	20:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:36
BLANK TEST	0.00	20:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:37
BLANK TEST	0.00	20:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:38
BLANK TEST	0.00	20:39

