

Infrared Instrument
Weekly Calibration Verification Test Record

Department: OCSO

Instrument Serial #: 123906

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
2/22/24	23430	3401	3400	3401	MP4189	.09	.09	.09	Dixon	NO
2/28/24	23430	3400	3401	3401	MP4189	.09	.09	.09	Osika	NO
3/1/24	23430	3401	3400	3401	MP4189	.10	.10	.10	Cusack	Yes
3/7/24	23430	3401	3401	3401	MP4189	.10	.10	.10	Dixon	NO
3/13/24	23430	3400	3401	3401	MP4189	.09	.09	.10	Osika	NO
3/21/24	23430	3401	3400	3401	MP4189	.09	.09	.10	Cusack	Yes
3/27/24	23430	3400	3400	3401	MP4189	.09	.09	.09	Dixon	NO
4/1/24	24040	3401	3401	3400	MP4189	.09	.10	.10	Osika	Yes
4/3/24	24040	3401	3401	3401	MP4189	.10	.10	.10	Dixon	NO
4/10/24	24040	3401	3401	3401	MP4189	.10	.10	.10	Osika	NO
4/17/24	24040	3401	3401	3401	MP4189	.09	.09	.10	MARCANO	NO
4/24/24	24040	3401	3401	3401	MP4189	.09	.09	.09	Osika	NO

* - Please enter a √ to denote simulator solution change taking place

CFO