

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 05/08/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:53
INTERNAL STANDARD	VERIFIED	01:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:53
BLANK TEST	0.00	01:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:54
BLANK TEST	0.00	01:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:55
BLANK TEST	0.00	01:56

