

SUPERVISOR TEST

LE13

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 05/16/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:04
INTERNAL STANDARD	VERIFIED	19:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:05
BLANK TEST	0.00	19:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:06
BLANK TEST	0.00	19:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:07
BLANK TEST	0.00	19:08

