

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 05/22/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:00
INTERNAL STANDARD	VERIFIED	05:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:00
BLANK TEST	0.00	05:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:02
BLANK TEST	0.00	05:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:03
BLANK TEST	0.00	05:04

