

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 06/19/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:53
INTERNAL STANDARD	VERIFIED	04:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:53
BLANK TEST	0.00	04:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:54
BLANK TEST	0.00	04:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:56
BLANK TEST	0.00	04:56

