

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 07/10/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:27
INTERNAL STANDARD	VERIFIED	22:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:27
BLANK TEST	0.00	22:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:28
BLANK TEST	0.00	22:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:30
BLANK TEST	0.00	22:30

