

SUPERVISOR TEST

LEB

ONEIDA COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 113606

Date: 05/30/2024

OPERATOR NAME:  
 JASON L OSIKA  
 Permit #: 35567  
 Operator Agency/Dept: ONEIDA CO. SO  
 Reference Standard #: 24040

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 17:19 |
| INTERNAL STANDARD       | VERIFIED | 17:19 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 17:20 |
| BLANK TEST              | 0.00     | 17:20 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 17:21 |
| BLANK TEST              | 0.00     | 17:22 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 17:22 |
| BLANK TEST              | 0.00     | 17:23 |

