

**Infrared Instrument
Weekly Calibration Verification Test Record**

Department: DESD

Instrument Serial #: 100197

| Test Date | Simulator Solution Lot # | Simulator Temperature | | | Sim. Serial# | Test Results | | | Breath Analysis Supervisor | Sim. Sol. Change * |
|-----------|--------------------------|-----------------------|-------|-------|---------------|--------------|-----|-------|----------------------------|--------------------|
| | | 1 | 2 | 3 | | 1 | 2 | 3 | | |
| 3/27/24 | 23430 | 34.03 | 34.03 | 34.03 | MP4192 | .09 | .09 | .09 | C. Brashear | Y |
| 3/27/24 | 23430 | 34.03 | 34.03 | 34.03 | MP4192 | .10 | .10 | .10 | C. Brashear | N |
| 4/1/24 | 24040 | 34.03 | 34.03 | 34.03 | MP4192 | .10 | .10 | .10 | Osika | Y |
| 4/4/24 | 24040 | 34.03 | 34.03 | 34.03 | MP4192 | .10 | .10 | .10 | Cusack | N |
| 4/10/24 | 24040 | 34.04 | 34.03 | 34.03 | MP4192 | .10 | .10 | .10 | Osika | N |
| 4/17/24 | 24040 | 34.04 | 34.04 | 34.04 | MP4192 | .10 | .10 | .10 | Brashear | N |
| 4/24/24 | 24040 | Error +3 | | | Detector Flow | | | Osika | N | |
| 5/8/24 | 24040 | Filter 3 won't zero | | | Zero | | | Osika | N | |
| 10-1-24 | 24270 | 34.03 | 34.03 | 34.03 | MP4192 | .10 | .10 | .10 | Eilers | Y |
| 10-9-24 | 24270 | 34.03 | 34.03 | 34.03 | MP4192 | .10 | .10 | .10 | Eilers | N |
| 10-16-24 | 24270 | 34.03 | 34.03 | 34.03 | MP4192 | .10 | .10 | .10 | Eilers | N |
| 10/23/24 | 24270 | 34.04 | 34.04 | 34.04 | MP4192 | .09 | .10 | .10 | Eilers | N |

* - Please enter a √ to denote simulator solution change taking place

Handwritten mark